

Mizuho Vascular Doppler Systems

Intraoperative essentials for neurosurgical and vascular procedures



Mizuho Vascular Doppler Systems

Intraoperative essentials for neurosurgical and vascular procedures

8 MHz System

2cm depth penetration depending on the density of surrounding tissue and bone

Indications

- **Locates carotid artery in transsphenoidal procedures**
- Allows technical verification of anastomoses and endarterectomies
- Supports assessment of flap viability
- Evaluates AV fistulas and EC-IC Bypass
- Verifies graft and shunt patency

Transceivers

- Available with A/C to D/C connection and battery only

Probes

- Packaged 4 in a box
- Sterile, Disposable
- Straight and Bayonet design
- 2mm or 3mm tip

Clinical Benefit Minimize the risk of internal carotid artery injury during transsphenoidal surgery



20 MHz System

1cm depth penetration depending on the density of surrounding tissue and bone

Indications

- **Determines completeness of aneurysm clipping**
- **Determines patency of parent vessel post aneurysm clipping**
- Locates feeder artery in AVM
- Locates vessels noted on angiogram
- Validates patency of microvascular anastomoses
- Confirms patency of major venous sinuses



Transceivers

- Available with A/C to D/C connection and battery only

Probes

- Packaged 4 in a box
- Sterile, Disposable
- Bayonet design only
- 2mm or 0.8mm tip



Clinical Benefit Quickly assess aneurysm clip placement and confirm aneurysm occlusion





Mizuho offers a portable, easy-to-use, and accurate modality for evaluating blood flow intraoperatively. Two systems operating at two different frequencies, 8MHz and 20MHz, provide audio blood flow detection at different depths of penetration. Each system comes with a set of transceivers and sterile disposable probes designed for specific surgical procedures.

Ordering Information



8 MHz System





	Product Code	Product Name	Unit	Size	Weight	Frequency	Power	Operating time
	0 7 - 1 5 0 - 0 8	8 MHz Doppler Transceiver with A/C	1	D 165mm × W 254mm × H 101mm	1.2kg	8MHz	8 AA Alkaline Batteries or A/C to D/C	4 hours, min.* (Standby mode, full volume)
	1 0 8 9 1 0	8 MHz Doppler Transceiver		D 105mm × W 95mm × H 135mm	0.6kg		8 AA Alkaline Batteries	10 hours, nom. (Standby mode, full volume)

	Product Code	Product Name	Pc	Tip length	Working length	Overall length	Shape of handle	Spec
	1 0 8 2 6 0	8 MHz Doppler Probe	4pcs /box	3mm	55mm	160mm	Straight	Sterile disposable
	1 3 8 2 1 0	8 MHz Kelly Endonasal Probe		2mm**	131mm	265mm	Bayoneted	

*During battery mode **Last 10mm of tip bent at 10 degrees

20 MHz System

	Product Code	Product Name	Unit	Size	Weight	Frequency	Power	Operating time
	0 7 - 1 5 0 - 2 0	20 MHz Doppler Transceiver with A/C	1	D 165mm × W 254mm × H 101mm	1.2kg	20MHz	8 AA Alkaline Batteries or A/C to D/C	4 hours, min.* (Standby mode, full volume)
	0 7 - 1 5 0 - 0 2	20 MHz Doppler Transceiver		D 105mm × W 95mm × H 135mm	0.6kg		8 AA Alkaline Batteries	10 hours, nom. (Standby mode, full volume)

	Product Code	Product Name	Pc	Tip length	Working length	Overall length	Shape of handle	Spec
	0 7 - 1 5 0 - 0 7	20 MHz Doppler Probe	4pcs /box	2mm	105mm	275mm	Bayoneted	Sterile disposable
	0 7 - 1 5 0 - 1 0	20 MHz Doppler Probe, Slim Handle			115mm	233mm		
	0 7 - 1 5 0 - 1 2	20 MHz Doppler Micro Probe		0.8mm	113mm	231mm		
	1 0 8 6 5 0	20 MHz Doppler Micro Probe, Slim Handle			74mm	200mm		

*During battery mode

Mizuho Vascular Doppler Systems



CE 0197

Mizuho Medical Co.,Ltd.

Hongo Shintoku Building 7F, 3-38-1
Hongo, Bunkyo-ku, Tokyo 113-0033, JAPAN

Phone +81-3-3815-7101
F a x +81-3-3815-1705
U R L www.mizuhomedical.co.jp
E-mail sales-dept@mizuhomedical.co.jp

Manufacturer
Vascular Technology
12 Murphy Dr.
Nashua, NH 03062, USA
Phone +1-800-550-0856



Caution

- Be sure not to use the product until you have carefully read the instruction manual and attached document and correctly understood their contents.
- Be sure to carry out without fail the periodic inspections and the pre- and post-operative inspections.
- For details about after-the-sale service be sure to inquire at our Sales Branches or the dealer from whom you have purchased this product.

* Under our policy, ongoing improvements and published specifications may be subject to change without prior notice.

* All published CG photos and illustrations have been from images.